

CERTIFICATE OF CALIBRATION Certificate No: 20211222235600195 ELC Manufacturer: **Product Type:** DC POWER SUPPLY Model: ALR3203 R/00195 Lot number: Serial No: **OF2356** 23±5°C **Humidity: Temperature:** < 75% **elc** procedure 9000 4 104-106 **Control and Tests performed:** Full procedure **Tests performed: Failed tests:** 0

Calibrated date: 2021-12-22

Output status:

This verification/calibration was performed with reference equipment connected to National Metrology Institutes, by competent personnel, following a manufacturer's procedure. This certificate shall not be reproduced, except in full, without the written approval of the calibration facility.

Within specification

REFERENCE EQUIPMENTS:

Description		Model	Serial No	Calibration due date
ELC Earth Continuity Tester		TC021	001	2021-04-30
SEFELEC Dielectric St	trength Tester	PC7P	135	2021-08-03
FLUKE Multimeter		FLUKE 45	6887016	2021-08-03
FLUKE Multimeter		FLUKE 45	6000037	2021-08-03
HEWLETT PACKARD Multimeter		34401A	US36083045	2021-08-03
ELC Shunt	•	SH40400	L00001	2021-08-03
ELC Electronic load		CE937	001	2021-04-30

Calibration Technician: AM

Checked by:

G. BONTEMPS

Note: The certificate applies only to this DC Power Supply. Without permission, you cannot copy a part of the certificate.