

## CERTIFICATE OF CALIBRATION

**Certificate No:** 20211222235600195  
**Manufacturer:** ELC  
**Product Type:** DC POWER SUPPLY  
**Model:** ALR3203  
**Serial No:** R/00195      **Lot number:** OF2356  
**Temperature:** 23±5°C      **Humidity:** < 75%

**Control and Tests performed:** elc procedure 9000 4 104-106  
**Tests performed:** Full procedure  
**Failed tests:** 0  
**Output status:** Within specification

**Calibrated date:** 2021-12-22

This verification/calibration was performed with reference equipment connected to National Metrology Institutes, by competent personnel, following a manufacturer's procedure.  
 This certificate shall not be reproduced, except in full, without the written approval of the calibration facility.

### REFERENCE EQUIPMENTS:

Description	Model	Serial No	Calibration due date
ELC Earth Continuity Tester	TC021	001	2021-04-30
SEFELEC Dielectric Strength Tester	PC7P	135	2021-08-03
FLUKE Multimeter	FLUKE 45	6887016	2021-08-03
FLUKE Multimeter	FLUKE 45	6000037	2021-08-03
HEWLETT PACKARD Multimeter	34401A	US36083045	2021-08-03
ELC Shunt	SH40400	L00001	2021-08-03
ELC Electronic load	CE937	001	2021-04-30

Calibration Technician: AM

Checked by: G. BONTEMPS



Note: The certificate applies only to this DC Power Supply.  
 Without permission, you cannot copy a part of the certificate.